Setting the Global Standard for Quality in Laboratory Testing P: 610.688.0100 Toll Free (US): 877.447.1888 F: 610.688.0700 E: customerservice@clsi.org

950 West Valley Road Suite 2500 Wayne, PA 19087



4 September 2015

To: Recipients of VET01S, 3rd ed.

From: Jennifer K. Adams, MT(ASCP), MSHA

Subject: Editorial Error in Table 4 and Editorial Update in Table 5 in CLSI Document VET01S

This notification is to inform you of two errors in Table 4 and two updates for consistency in Table 5 in CLSI document VET01S, 3rd ed., Performance Standards for Antimicrobial Disk and Dilution Susceptibility Tests for Bacteria Isolated From Animals.

Table 4: The QC ranges for the following drugs/antimicrobial agents were incorrect:

- Escherichia coli ATCC® 25922 Ampicillin
- Pseudomonas aeruginosa ATCC® 27853 Gentamicin

The corrected ranges are highlighted in the excerpted rows from Table 4, below:

Table 4. Acceptable Quality Control Ranges of Antimicrobial Disk Susceptibility Test Zone Diameters (mm) for Reference Strains on Mueller-Hinton Agar (Except Where Noted)

Antimicrobial Agent	Disk Content	Escherichia coli ATCC®a 25922	Staphylococcus aureus ATCC® 25923	Pseudomonas aeruginosa ATCC® 27853	Streptococcus pneumoniae ^b ATCC® 49619
Antimicrobiai Agent Ampicillin	10 µg	25922 15–22	27–35	ATCC 2/855	30–36
Gentamicin	10 μg	19–26	19–27	17–23	-

Table 5: The QC dilution ranges for rifampin with *Staphylococcus aureus* ATCC® 29213 and *Streptococcus* pneumoniae ATCC® 49619 were updated for consistency with CLSI document M100. The edits, which do not represent a substantive change, are highlighted in the excerpted row from Table 5, below:

Table 5. Acceptable Quality Control Ranges of Minimal Inhibitory Concentrations for Broth

Microdilution¹ (ug/mL) for Reference Strains^a

	Staphylococcus	Enterococcus	Escherichia	Pseudomonas	Streptococcus			
	aureus	faecalis	coli	aeruginosa	pneumoniae ^c			
Antimicrobial Agent	ATCC® 29213	ATCC® 29212	ATCC® 25922	ATCC® 27853	ATCC® 49619			
Rifampin	0.004-0.01 <mark>5</mark>	0.5-4	4–16	16–64	0.01 <mark>5</mark> –0.06			

If you require any additional clarification regarding this correction, please contact CLSI Customer Service (customerservice@clsi.org).

We appreciate your commitment to CLSI, and regret any inconvenience.