Setting the **Global** Standard for Quality in Laboratory Testing

P: +1.610.688.0100 **Toll Free (US):** 877.447.1888 Suite 2500 F: +1.610.688.0700 E: customerservice@clsi.org

950 West Valley Road Wayne, PA 19087 USA



10 April 2018

To: Recipients of M100, 28th ed.

From: Jennifer K. Adams, MT(ASCP), MSHA

Vice President, Standards and Quality

Subject: Correction

This notification is to inform you of a correction made to CLSI document M100, Performance Standards for Antimicrobial Susceptibility Testing, 28th ed. The correction is described below and shown as highlighted and stricken text in the table excerpt.

Table 4A-2, Disk Diffusion QC Ranges for Nonfastidious Organisms and B-Lactam Combination Agents:

The disk content provided in Table 4A-2 for ceftazidime-avibactam is listed incorrectly as " $30/15 \mu q$."

In Table 4A-2, the disk content for ceftazidime-avibactam has been corrected to read "30/20 μq."

The Table 4A-2 correction is shown in the table excerpt below.

Table 4A-2. Disk Diffusion QC Ranges for Nonfastidious Organisms and B-Lactam Combination Agents

		QC Organisms and Characteristics					
		Escherichia			Escherichia	Klebsiella	
		coli	Pseudomonas	Staphylococcus	coli	pneumoniae	
		ATCC®a	aeruginosa	aureus	ATCC ^{®b,c}	ATCC®	
		25922	ATCC® 27853	ATCC® 25923	35218	700603 ^{b,c}	
						SHV-18	
						OXA-2	
						Mutations in	
		0 1		B-lactamase		OmpK35 and	
		B-lactamase negative	Inducible AmpC	negative, <i>mecA</i> negative	TEM-1	OmpK37 TEM-1	
Antimicrobial	Disk	Hegative	iliducible Allipc	Hegalive	LM-	L/M-1	VIIIIIIIIIIIIIII
Agent	Content	MIC QC ranges, mm					
Ceftazidime-	30/ 15	27-35	25-31	16-22	28-35	21-27 ^e	
avibactam	<mark>20</mark> μg	27 33	25 51	10 22	20 33	21-21	

If you require any additional clarification regarding these corrections, please contact CLSI Customer Service (customerservice@clsi.org).

We appreciate your commitment to CLSI and regret any inconvenience.