

24 May 2023

To: Recipients of M24S-Ed2

From: Jennifer K. Adams, MLS(ASCP), MSHA
Vice President, Standards and Quality

Subject: Correction

This notice is intended to inform users of a correction made to CLSI document M24S, *Performance Standards for Susceptibility Testing of Mycobacteria, Nocardia spp., and Other Aerobic Actinomycetes*, 2nd ed. The correction is described below and shown as stricken and highlighted text in the table excerpt.

Table 10. MIC QC Ranges When Testing MTBC Using *Mycobacterium tuberculosis* ATCC® 27294 (H37Rv) Tested in Middlebrook 7H9 Medium Supplemented With OADC

Antimicrobial Agent	MIC QC Ranges, µg/mL	
	Lyophilized MIC Panels	Frozen MIC Panels (prepared according to CLSI documents M07 and M24)
Rifampin	≤0.12	0.06-0.25 0.03-0.25

If you require any additional clarification regarding these corrections, please contact CLSI Customer Service (customerservice@clsi.org).

We appreciate your commitment to CLSI and regret any inconvenience.