



7 April 2022

- To: Recipients of H62, 1st ed.
- From: Jennifer K. Adams, MT(ASCP), MSHA Vice President, Standards and Quality
- Subject: Combined corrections

This notice is intended to inform users of corrections made to CLSI document H62, *Validation of Assays Performed by Flow Cytometry*, 1st ed. The corrections are described below and shown in red in Figure 22C and as highlighted and/or stricken text in the excerpts.

Correction: 7 April 2022

Figure 22C. Analytical Validation:

Figure 22C, Analytical Validation, has been corrected to show Run 3 is Operator 1 on Instrument $\frac{4}{2}$ and Run 4 is Operator 2 on Instrument $\frac{2}{2}$ 1.



Clinical and Laboratory Standards Institute PO Box 334, Malvern, PA 19355 USA | www.clsi.org P: +1.610.688.0100 Toll Free (US): 877.447.1888 F: +1.610.688.0700 E: customerservice@clsi.org

Correction: 17 December 2021

Foreword:

The proprietary names "Triton^M X-100" and "Tween^M 20" were not identified as trademarks. The trademark disclaimer has been corrected to read "Use of Alexa Fluor 488[®] (succinimidyl ester free acid), Brilliant^M Blue, Brilliant Violet^M, PE-Texas Red[®] (phycoerythrin-sulforhodamine 101 sulfonyl chloride), and SRM[®] 1934, Triton^M X-100 (alkyl polyglucoside), and Tween^M 20 (polysorbate 20) in this guideline is not an endorsement on the part of CLSI. With each use of the trade name, "or the equivalent" is added to indicate that this guideline also applies to any equivalent products."

Appendix B. Critical Reagents Beyond Antibodies:

The text in B5, Detergents and Solvents, has been corrected to read "Detergents such as saponin, NP-40, Triton^m X-100 (alkyl polyglucoside [or equivalent]), or Tween^m 20 (polysorbate 20 [or equivalent]) may be used for samples that require lysing." In addition to the text correction, the following footnotes have been added: "^[a] Triton^m X-100 is a registered trademark of the Dow Chemical Company" and "^[b] Tween^m 20 is a registered trademark of Croda Americas."

If you require any additional clarification regarding these corrections, please contact CLSI Customer Service (<u>customerservice@clsi.org</u>).

We appreciate your commitment to CLSI and regret any inconvenience.