

6 May 2025

To: Recipients of CLSI M07-Ed12

From: Jennifer K. Adams, MLS(ASCP), MSHA
Vice President, Standards and Quality

Subject: Correction

This notice is intended to inform users of a correction made to CLSI M07, *Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria That Grow Aerobically*, 12th ed. The correction is described below and shown as highlighted and/or stricken text in the table excerpt.

A3.3 Iron-Depleted Cation-Adjusted Mueller-Hinton Broth:

The measurement unit for Zn^{++} is listed incorrectly as “mmol.” The measurement unit has been corrected to read “ μmol .”

Step	Action	Comment
1	Dissolve 0.29 g $\text{ZnSO}_4 \cdot 7\text{H}_2\text{O}$ in 100 mL deionized water.	<ul style="list-style-type: none">This solution contains 0.65 mg Zn^{++}/mL (10 mmol μmol Zn^{++}/mL).Verify that the deionized water has an iron content of $\leq 0.03 \text{ mg/L}$.

If you require any additional clarification regarding this correction, please contact CLSI Customer Service (customerservice@clsi.org).